

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1327 SAS No.: Q1327 SDG NO.: Q1327

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 02/13/2025

Lab File ID: BN036459.D Time Analyzed: 09:48

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	2295	7.746	6046	10.54	4260	14.39
UPPER LIMIT	4590	8.246	12092	11.041	8520	14.888
LOWER LIMIT	1147.5	7.246	3023	10.041	2130	13.888
EPA SAMPLE NO.						
01 VPB192-HYD-20250204RE	2222	7.75	5257	10.54	3408	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.