

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1327

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX044897.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0211WBSD01	Chloromethane	20	18.7	ug/L	94	5		50	139	20
	Vinyl chloride	20	19.2	ug/L	96	6		58	137	20
	Bromomethane	20	20.1	ug/L	101	6		53	141	20
	Chloroethane	20	18.9	ug/L	95	17		60	138	20
	Trichlorofluoromethane	20	19.5	ug/L	98	6		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	20.3	ug/L	102	9		70	136	20
	1,1-Dichloroethene	20	19.3	ug/L	97	7		71	131	20
	Acetone	100	100	ug/L	100	13		39	160	20
	Carbon disulfide	20	18.6	ug/L	93	8		64	133	20
	Methyl tert-butyl Ether	20	20.2	ug/L	101	13		71	124	20
	Methylene Chloride	20	19.6	ug/L	98	10		74	124	20
	trans-1,2-Dichloroethene	20	19.7	ug/L	99	10		75	124	20
	1,1-Dichloroethane	20	19.6	ug/L	98	7		77	125	20
	2-Butanone	100	100	ug/L	100	11		56	143	20
	Carbon Tetrachloride	20	20.0	ug/L	100	3		72	136	20
	cis-1,2-Dichloroethene	20	20.0	ug/L	100	7		78	123	20
	Chloroform	20	19.7	ug/L	99	6		79	124	20
	1,1,1-Trichloroethane	20	19.6	ug/L	98	7		74	131	20
	Methylcyclohexane	20	20.8	ug/L	104	2		72	132	20
	Benzene	20	20.2	ug/L	101	5		79	120	20
	1,2-Dichloroethane	20	21.2	ug/L	106	7		73	128	20
	Trichloroethene	20	19.7	ug/L	99	3		79	123	20
	1,2-Dichloropropane	20	20.4	ug/L	102	6		78	122	20
	Bromodichloromethane	20	20.8	ug/L	104	6		79	125	20
	4-Methyl-2-Pentanone	100	110	ug/L	110	11		67	130	20
	Toluene	20	20.9	ug/L	104	6		80	121	20
	t-1,3-Dichloropropene	20	20.0	ug/L	100	7		73	127	20
	cis-1,3-Dichloropropene	20	20.5	ug/L	103	7		75	124	20
	1,1,2-Trichloroethane	20	21.3	ug/L	106	7		80	119	20
	2-Hexanone	100	110	ug/L	110	10		57	139	20
	Dibromochloromethane	20	21.0	ug/L	105	10		74	126	20
	Tetrachloroethene	20	20.0	ug/L	100	6		74	129	20
	Chlorobenzene	20	20.3	ug/L	102	7		82	118	20
	Ethyl Benzene	20	20.5	ug/L	103	7		79	121	20
	m/p-Xylenes	40	41.7	ug/L	104	7		80	121	20
	o-Xylene	20	20.5	ug/L	103	7		78	122	20
	Styrene	20	20.9	ug/L	104	7		78	123	20
	Bromoform	20	20.4	ug/L	102	6		66	130	20
	Isopropylbenzene	20	20.5	ug/L	103	11		72	131	20
	1,1,2,2-Tetrachloroethane	20	20.4	ug/L	102	9		71	121	20
	1,3-Dichlorobenzene	20	20.1	ug/L	101	5		80	119	20
	1,4-Dichlorobenzene	20	20.0	ug/L	100	6		79	118	20
	1,2-Dichlorobenzene	20	20.8	ug/L	104	7		80	119	20