

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1327 SAS No.: Q1327 SDG NO.: Q1327
Lab File ID: VX044893.D Date Analyzed: 02/11/2025
Instrument ID: MSVOA_X Time Analyzed: 10:09
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	125474	5.54	221806	6.75	200705	10.05
UPPER LIMIT	250948	6.044	443612	7.251	401410	10.549
LOWER LIMIT	62737	5.044	110903	6.251	100353	9.549
EPA SAMPLE NO.						
VPB192-HYD-20250204	94029	5.54	191186	6.76	175343	10.05
BP-VPB-192-TB-20250203	91689	5.54	184061	6.76	168042	10.05
BP-VPB-192-DUP-20250204	91782	5.54	182308	6.76	169727	10.05
BP-VPB-192-GW-520-522	88393	5.55	180540	6.76	161074	10.05
BP-VPB-192-GW-540-542	86070	5.54	174840	6.76	160189	10.05
BP-VPB-192-GW-560-562	90647	5.54	181380	6.76	162703	10.05
BP-VPB-192-GW-580-582	91582	5.55	182202	6.76	164292	10.05
BP-VPB-192-GW-600-602	93253	5.54	190982	6.76	174939	10.05
BP-VPB-192-GW-625-627	90365	5.55	184452	6.76	165045	10.05
VX0211WBL01	98496	5.54	199075	6.76	183423	10.05
VX0211WBS01	122388	5.54	213544	6.76	190238	10.05
VX0211WBSD01	116406	5.54	211404	6.76	186335	10.05

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.