

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1327 SAS No.: Q1327 SDG NO.: Q1327
Lab File ID: VX044893.D Date Analyzed: 02/11/2025
Instrument ID: MSVOA_X Time Analyzed: 10:09
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	89456	12.018				
UPPER LIMIT	178912	12.518				
LOWER LIMIT	44728	11.518				
EPA SAMPLE NO.						
VPB192-HYD-20250204	74493	12.02				
BP-VPB-192-TB-20250203	72398	12.02				
BP-VPB-192-DUP-20250204	71217	12.02				
BP-VPB-192-GW-520-522	67868	12.02				
BP-VPB-192-GW-540-542	67837	12.02				
BP-VPB-192-GW-560-562	65512	12.02				
BP-VPB-192-GW-580-582	67262	12.02				
BP-VPB-192-GW-600-602	75627	12.02				
BP-VPB-192-GW-625-627	71791	12.02				
VX0211WBL01	79632	12.02				
VX0211WBS01	87843	12.02				
VX0211WBSD01	83644	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.