

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1328 SAS No.: Q1328 SDG NO.: Q1328

Continuing Calib Date: 02/11/2025 Initial Calibration Date(s): 02/03/2025 02/04/2025

Continuing Calib Time: 17:19 Initial Calibration Time(s): 15:58 02:03

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.50	4.50	4.43	4.57	0.00
Aroclor-1016-2	(2)	4.52	4.52	4.45	4.59	0.00
Aroclor-1016-3	(3)	4.58	4.58	4.51	4.65	0.00
Aroclor-1016-4	(4)	4.67	4.67	4.60	4.74	0.00
Aroclor-1016-5	(5)	4.95	4.95	4.88	5.02	0.00
Aroclor-1260-1	(1)	6.04	6.04	5.97	6.11	0.00
Aroclor-1260-2	(2)	6.30	6.30	6.23	6.37	0.00
Aroclor-1260-3	(3)	6.65	6.65	6.58	6.72	0.00
Aroclor-1260-4	(4)	6.86	6.86	6.79	6.93	0.00
Aroclor-1260-5	(5)	7.17	7.17	7.10	7.24	0.00
TCX		3.42	3.42	3.37	3.47	0.00
DCB		8.53	8.53	8.43	8.63	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1328 SAS No.: Q1328 SDG NO.: Q1328

Continuing Calib Date: 02/11/2025 Initial Calibration Date(s): 02/03/2025 02/04/2025

Continuing Calib Time: 17:19 Initial Calibration Time(s): 15:58 02:03

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	3.79	3.79	3.72	3.86	0.00
Aroclor-1016-2	(2)	3.81	3.81	3.74	3.88	0.00
Aroclor-1016-3	(3)	3.97	3.97	3.90	4.04	0.00
Aroclor-1016-4	(4)	4.01	4.01	3.94	4.08	0.00
Aroclor-1016-5	(5)	4.21	4.21	4.14	4.28	0.00
Aroclor-1260-1	(1)	5.21	5.21	5.14	5.28	0.01
Aroclor-1260-2	(2)	5.39	5.39	5.32	5.46	0.00
Aroclor-1260-3	(3)	5.54	5.54	5.47	5.61	0.00
Aroclor-1260-4	(4)	6.00	6.00	5.93	6.07	0.00
Aroclor-1260-5	(5)	6.24	6.24	6.17	6.31	0.00
TCX		2.79	2.79	2.74	2.84	0.00
DCB		7.56	7.56	7.46	7.66	0.00