

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1328 SAS No.: Q1328 SDG NO.: Q1328

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/03/2025 02/03/2025

Client Sample No.: CCAL01 Date Analyzed: 02/11/2025

Lab Sample No.: AR1660CCC400 Data File : PQ069975.D Time Analyzed: 17:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.499	4.400	4.600	246428000.000	230460000.000	-6.5
Aroclor-1016-2	4.518	4.418	4.618	401073200.000	377784000.000	-5.8
Aroclor-1016-3	4.576	4.476	4.676	254289000.000	243932000.000	-4.1
Aroclor-1016-4	4.666	4.565	4.765	205550800.000	196814000.000	-4.3
Aroclor-1016-5	4.951	4.850	5.050	202136200.000	194915000.000	-3.6
Aroclor-1260-1	6.042	5.942	6.142	346327400.000	331040000.000	-4.4
Aroclor-1260-2	6.298	6.198	6.398	394244700.000	362855000.000	-8.0
Aroclor-1260-3	6.650	6.549	6.749	307979900.000	287401000.000	-6.7
Aroclor-1260-4	6.864	6.764	6.964	334749900.000	312932000.000	-6.5
Aroclor-1260-5	7.170	7.071	7.271	624266900.000	565993000.000	-9.3
DCB	8.527	8.427	8.627	3793119000.000	3548280000.000	-6.5
TCX	3.416	3.316	3.516	8953265000.000	8492160000.000	-5.2

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1328 SAS No.: Q1328 SDG NO.: Q1328

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/03/2025 02/03/2025

Client Sample No.: CCAL01 Date Analyzed: 02/11/2025

Lab Sample No.: AR1660CCC400 Data File : PQ069975.D Time Analyzed: 17:19

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.789	3.689	3.889	241461400.000	219367000.000	-9.2
Aroclor-1016-2	3.805	3.705	3.905	372292200.000	345706000.000	-7.1
Aroclor-1016-3	3.966	3.866	4.066	209142600.000	195612000.000	-6.5
Aroclor-1016-4	4.014	3.914	4.114	183680500.000	173829000.000	-5.4
Aroclor-1016-5	4.210	4.110	4.310	220642100.000	207234000.000	-6.1
Aroclor-1260-1	5.205	5.106	5.306	400982800.000	371887000.000	-7.3
Aroclor-1260-2	5.394	5.294	5.494	475736000.000	437324000.000	-8.1
Aroclor-1260-3	5.536	5.436	5.636	445931200.000	408870000.000	-8.3
Aroclor-1260-4	5.998	5.898	6.098	389237200.000	364279000.000	-6.4
Aroclor-1260-5	6.243	6.143	6.343	865239900.000	798723000.000	-7.7
DCB	7.559	7.459	7.659	6072876000.000	5623070000.000	-7.4
TCX	2.786	2.685	2.885	7930420000.000	7396960000.000	-6.8