

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1328 SAS No.: Q1328 SDG NO.: Q1328

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/03/2025 02/03/2025

Client Sample No.: CCAL02 Date Analyzed: 02/11/2025

Lab Sample No.: AR1660CCC400 Data File : PQ069988.D Time Analyzed: 20:40

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.497	4.400	4.600	246428000.000	230093000.000	-6.6
Aroclor-1016-2	4.516	4.418	4.618	401073200.000	366001000.000	-8.7
Aroclor-1016-3	4.574	4.476	4.676	254289000.000	233721000.000	-8.1
Aroclor-1016-4	4.664	4.565	4.765	205550800.000	190204000.000	-7.5
Aroclor-1016-5	4.948	4.850	5.050	202136200.000	187533000.000	-7.2
Aroclor-1260-1	6.039	5.942	6.142	346327400.000	322307000.000	-6.9
Aroclor-1260-2	6.296	6.198	6.398	394244700.000	363108000.000	-7.9
Aroclor-1260-3	6.647	6.549	6.749	307979900.000	283006000.000	-8.1
Aroclor-1260-4	6.862	6.764	6.964	334749900.000	307837000.000	-8.0
Aroclor-1260-5	7.168	7.071	7.271	624266900.000	570161000.000	-8.7
DCB	8.524	8.427	8.627	3793119000.000	3535120000.000	-6.8
TCX	3.414	3.316	3.516	8953265000.000	8177130000.000	-8.7

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1328 SAS No.: Q1328 SDG NO.: Q1328

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/03/2025 02/03/2025

Client Sample No.: CCAL02 Date Analyzed: 02/11/2025

Lab Sample No.: AR1660CCC400 Data File : PQ069988.D Time Analyzed: 20:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	3.789	3.689	3.889	241461400.000	203752000.000	-15.6
Aroclor-1016-2	3.804	3.705	3.905	372292200.000	313346000.000	-15.8
Aroclor-1016-3	3.966	3.866	4.066	209142600.000	188702000.000	-9.8
Aroclor-1016-4	4.014	3.914	4.114	183680500.000	164838000.000	-10.3
Aroclor-1016-5	4.210	4.110	4.310	220642100.000	201757000.000	-8.6
Aroclor-1260-1	5.205	5.106	5.306	400982800.000	357391000.000	-10.9
Aroclor-1260-2	5.394	5.294	5.494	475736000.000	425759000.000	-10.5
Aroclor-1260-3	5.536	5.436	5.636	445931200.000	398851000.000	-10.6
Aroclor-1260-4	5.997	5.898	6.098	389237200.000	351964000.000	-9.6
Aroclor-1260-5	6.242	6.143	6.343	865239900.000	784847000.000	-9.3
DCB	7.558	7.459	7.659	6072876000.000	5418900000.000	-10.8
TCX	2.786	2.685	2.885	7930420000.000	7051410000.000	-11.1