

Surrogate Summary

SDG No.: Q1328

Client: Chemtech Consulting Group

Analytical Method: SFAM_PCB

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PQ069912.D	PIBLK-PQ069912.D	TCX	1	20	0.19	97		30	150
		DCB	1	40	0.41	102		30	150
		TCX	2	20	0.21	103		30	150
		DCB	2	40	0.41	102		30	150
I.BLK-PQ069974.D	PIBLK-PQ069974.D	TCX	1	20	0.20	101		30	150
		DCB	1	40	0.42	104		30	150
		TCX	2	20	0.21	103		30	150
		DCB	2	40	0.40	101		30	150
Q1328-13	PCB-GPC2-BLANK	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
Q1328-14	PCB-GPC2-BLANK-SPIKE	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
I.BLK-PQ069987.D	PIBLK-PQ069987.D	TCX	1	20	0.19	95		30	150
		DCB	1	40	0.40	101		30	150
		TCX	2	20	0.20	99		30	150
		DCB	2	40	0.39	97		30	150