

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1328 SAS No.: Q1328 SDG NO.: Q1328

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/16/2025 01/16/2025

Client Sample No.: CCAL01 Date Analyzed: 02/12/2025

Lab Sample No.: INDA315 Data File : PD087860.D Time Analyzed: 13:20

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	6.711	6.641	6.781	2749490.000	2919390.000	6.2
4,4'-DDT	7.027	6.957	7.097	2898230.000	2550180.000	-12.0
alpha-BHC	4.005	3.955	4.055	4154340.000	4736570.000	14.0
DCB	9.082	8.982	9.182	3343300.000	3228600.000	-3.4
Dieldrin	6.353	6.283	6.423	3812360.000	3995740.000	4.8
Endosulfan I	6.081	6.011	6.151	3483940.000	3671150.000	5.4
Endrin	6.581	6.511	6.651	3245680.000	3013440.000	-7.2
gamma-BHC (Lindane)	4.336	4.286	4.386	4022600.000	4369180.000	8.6
Heptachlor	4.936	4.886	4.986	4087560.000	4363480.000	6.8
Methoxychlor	7.499	7.429	7.569	1584350.000	1414180.000	-10.7
TCX	3.555	3.505	3.605	2037980.000	2349570.000	15.3

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1328 SAS No.: Q1328 SDG NO.: Q1328

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/16/2025 01/16/2025

Client Sample No.: CCAL01 Date Analyzed: 02/12/2025

Lab Sample No.: INDA315 Data File : PD087860.D Time Analyzed: 13:20

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
4,4'-DDD	5.934	5.864	6.004	13369700.000	15970000.000	19.4
4,4'-DDT	6.189	6.119	6.259	13780800.000	13914900.000	1.0
alpha-BHC	3.395	3.345	3.445	19801700.000	23522900.000	18.8
DCB	8.079	7.979	8.179	14484400.000	15141900.000	4.5
Dieldrin	5.517	5.447	5.587	16690700.000	19084300.000	14.3
Endosulfan I	5.251	5.181	5.321	15872300.000	17871700.000	12.6
Endrin	5.793	5.723	5.863	14942300.000	15848100.000	6.1
gamma-BHC (Lindane)	3.731	3.681	3.781	18316800.000	21459300.000	17.2
Heptachlor	4.086	4.036	4.136	18664300.000	21226700.000	13.7
Methoxychlor	6.760	6.690	6.830	6159300.000	6346980.000	3.0
TCX	2.881	2.831	2.931	12367800.000	14896700.000	20.4