

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1328 SAS No.: Q1328 SDG NO.: Q1328

EPA Sample No.: SSTD020618 Date Analyzed: 01/29/2025

Lab File ID: BP023786.D Time Analyzed: 14:38

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	487689	7.787	2130100	10.56	1425430	14.42
UPPER LIMIT	975378	8.287	4260200	11.063	2850860	14.916
LOWER LIMIT	243845	7.287	1065050	10.063	712715	13.916
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	356514	7.76	1536300	10.53	978847	14.38

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.