

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1328 SAS No.: Q1328 SDG NO.: Q1328

EPA Sample No.: SSTD020657 Date Analyzed: 02/11/2025

Lab File ID: BP024032.D Time Analyzed: 10:16

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	414644	7.758	1750430	10.53	1093690	14.38
UPPER LIMIT	829288	8.258	3500860	11.028	2187380	14.881
LOWER LIMIT	207322	7.258	875215	10.028	546845	13.881
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	356514	7.76	1536300	10.53	978847	14.38

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.