

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1328 SAS No.: Q1328 SDG NO.: Q1328

EPA Sample No.: SSTD020657 Date Analyzed: 02/11/2025

Lab File ID: BP024032.D Time Analyzed: 10:16

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	2137420	17.169	2177180	21.604	2289570	24.945
UPPER LIMIT	4274840	17.669	4354360	22.104	4579140	25.445
LOWER LIMIT	1068710	16.669	1088590	21.104	1144790	24.445
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	1966560	17.16	1799470	21.60	1786510	24.95

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.