

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CHEM02

Lab Code: CHEM Case No.: Q1328 SAS No.: Q1328 SDG No.: Q1328

Instrument ID: MSVOA_N Calibration Date/Time: 02/10/2025 12:12

Lab File ID: VN085718.D Init. Calib. Date(s): 01/14/2025 01/14/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 14:56 17:19

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
p-Isopropyltoluene	50.000	53.56	-7.127998	20