

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1328 SAS No.: Q1328 SDG NO.: Q1328
 Lab File ID: VN085718.D Date Analyzed: 02/10/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:12
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	241607	8.22	411898	9.09	357680	11.87
UPPER LIMIT	483214	8.718	823796	9.594	715360	12.365
LOWER LIMIT	120804	7.718	205949	8.594	178840	11.365
EPA SAMPLE NO.						
Storage-Blank-SOIL-REF-6	220063	8.22	405237	9.10	352429	11.87
Storage-Blank-WATER-Ref-5	229831	8.22	425415	9.10	373084	11.87
Storage-Blank-WATER-REF-4	225145	8.22	415405	9.10	357489	11.87
Storage-Blank-SAMLE-MANAGEMENT	215052	8.22	406991	9.10	354283	11.87
VN0210WBL01	252400	8.22	469407	9.10	394938	11.87
VN0210WBS01	320498	8.22	543900	9.10	477598	11.87
VN0210WBSD01	231142	8.22	403589	9.10	358601	11.87

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.