

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1328 SAS No.: Q1328 SDG NO.: Q1328
 Lab File ID: VN085718.D Date Analyzed: 02/10/2025
 Instrument ID: MSVOA_N Time Analyzed: 12:12
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	178725	13.788				
UPPER LIMIT	357450	14.288				
LOWER LIMIT	89362.5	13.288				
EPA SAMPLE NO.						
Storage-Blank-SOIL-REF-6	133160	13.79				
Storage-Blank-WATER-Ref-5	146687	13.79				
Storage-Blank-WATER-REF-4	132906	13.79				
Storage-Blank-SAMPLE-MANAGEMENT	136368	13.79				
VN0210WBL01	156176	13.79				
VN0210WBS01	232853	13.79				
VN0210WBSD01	166903	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.