DATA OF KNOWN QUALITY CONFORMANCE/NON-CONFORMANCE SUMMARY QUESTIONNAIRE

Laboratory Name :		CHEMTECH	Client :	G Environmental					
Project Location : Newark, NJ		Newark, NJ	Project Number :	- Power					
Laboratory Sample ID(s): Q1331			Sampling Date(s):	2/06/2025					
List DKQP Methods Used (e.g., 8260,8270, et Cetra) 8260-Low,8270E,SOP									
1	specified QA/Q explain any crit	tical method referenced in this labo C performance criteria followed, in eria falling outside of acceptable g Known Quality performance stand	cluding the requirement to uidelines, as specified in the		V	Yes		No	
1A	Were the method specified handling, preservation, and holding time requirements met?					Yes		No	
1B	EPH Method: Was the EPH method conducted without significant modifications (see Section 11.3 of respective DKQ methods)					Yes		No	□ N/A
2	Were all samples received by the laboratory in a condition consistent with that described on the associated chain-of-custody document(s)?				V	Yes		No	
3	Were samples received at an appropriate temperature (4±2° C)?				V	Yes		No	□ N/A
4	Were all QA/QC performance criteria specified in the NJDEP DKQP standards achieved?					Yes	V	No	
5	1 '	ng limits specified or referenced on to the laboratory prior to sample re	-		V	Yes		No	
	b)Were these re	eporting limits met?			V	Yes		No	□ N/A
6	results reporte	tical method referenced in this labor d for all constituents identified in the e DKQP documents and/or site-sp	e method-specific analyte lists		V	Yes		No	
7	Are project-spe	cific matrix spikes and/or laborator	y duplicates included in this dat	a set?	V	Yes		No	

Notes: For all questions to which the response was "No" (with the exception of question #7), additional information should be provided in an attached narrative. If the answer to question #1, #1A, or #1B is "No", the data package does not meet the requirements for "Data of Known Quality."