

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1331 SAS No.: Q1331 SDG NO.: Q1331

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/11/2025

Lab File ID: BF141551.D Time Analyzed: 10:28

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	100982	6.793	399334	8.08	220290	9.83
UPPER LIMIT	201964	7.293	798668	8.575	440580	10.328
LOWER LIMIT	50491	6.293	199667	7.575	110145	9.328
EPA SAMPLE NO.						
01 PB166641BL	97044	6.79	390720	8.07	219692	9.82
02 PB166641BS	98554	6.79	391258	8.07	220324	9.83
03 PB166641BSD	100463	6.79	407334	8.07	229450	9.83

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.