

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1331 SAS No.: Q1331 SDG NO.: Q1331

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/14/2025

Lab File ID: BF141639.D Time Analyzed: 15:04

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	108299	6.787	422251	8.07	233877	9.82
UPPER LIMIT	216598	7.287	844502	8.569	467754	10.322
LOWER LIMIT	54149.5	6.287	211126	7.569	116939	9.322
EPA SAMPLE NO.						
01 MW1RDL	69665	6.79	254515	8.08	116077 *	9.85

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.