

Surrogate Summary

SDG No.: **Q1332**

Client: **G Environmental**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1332-01	MW1	1,2-Dichloroethane-d4	50	45.6	91	70 (74)	130 (125)
		Dibromofluoromethane	50	48.7	97	70 (75)	130 (124)
		Toluene-d8	50	52.9	106	70 (86)	130 (113)
		4-Bromofluorobenzene	50	55.5	111	70 (77)	130 (121)
Q1332-01DL	MW1DL	1,2-Dichloroethane-d4	50	52.9	106	70 (74)	130 (125)
		Dibromofluoromethane	50	50.0	100	70 (75)	130 (124)
		Toluene-d8	50	50.4	101	70 (86)	130 (113)
		4-Bromofluorobenzene	50	54.1	108	70 (77)	130 (121)
VN0210WBL01	VN0210WBL01	1,2-Dichloroethane-d4	50	47.8	96	70 (74)	130 (125)
		Dibromofluoromethane	50	50.2	100	70 (75)	130 (124)
		Toluene-d8	50	48.3	97	70 (86)	130 (113)
		4-Bromofluorobenzene	50	44.2	88	70 (77)	130 (121)
VN0210WBS01	VN0210WBS01	1,2-Dichloroethane-d4	50	39.3	79	70 (74)	130 (125)
		Dibromofluoromethane	50	44.5	89	70 (75)	130 (124)
		Toluene-d8	50	44.5	89	70 (86)	130 (113)
		4-Bromofluorobenzene	50	46.6	93	70 (77)	130 (121)
VN0210WBSD01	VN0210WBSD01	1,2-Dichloroethane-d4	50	49.9	100	70 (74)	130 (125)
		Dibromofluoromethane	50	54.4	109	70 (75)	130 (124)
		Toluene-d8	50	54.6	109	70 (86)	130 (113)
		4-Bromofluorobenzene	50	57.5	115	70 (77)	130 (121)
VX0211WBL01	VX0211WBL01	1,2-Dichloroethane-d4	50	53.5	107	70 (74)	130 (125)
		Dibromofluoromethane	50	51.0	102	70 (75)	130 (124)
		Toluene-d8	50	50.3	100	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.4	105	70 (77)	130 (121)
VX0211WBS01	VX0211WBS01	1,2-Dichloroethane-d4	50	43.8	88	70 (74)	130 (125)
		Dibromofluoromethane	50	45.5	91	70 (75)	130 (124)
		Toluene-d8	50	46.2	92	70 (86)	130 (113)
		4-Bromofluorobenzene	50	47.4	95	70 (77)	130 (121)

