

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1332 SAS No.: Q1332 SDG NO.: Q1332
Lab File ID: VN085718.D Date Analyzed: 02/10/2025
Instrument ID: MSVOA_N Time Analyzed: 12:12
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	241607	8.22	411898	9.09	357680	11.87
UPPER LIMIT	483214	8.718	823796	9.594	715360	12.365
LOWER LIMIT	120804	7.718	205949	8.594	178840	11.365
EPA SAMPLE NO.						
MW1	247634	8.22	449721	9.10	407986	11.87
VN0210WBL01	252400	8.22	469407	9.10	394938	11.87
VN0210WBS01	320498	8.22	543900	9.10	477598	11.87
VN0210WBSD01	231142	8.22	403589	9.10	358601	11.87

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.