

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51900 MA No.: _____ SDG No.: ME29C5
Instrument ID: KONELAB Start Date: 02/13/2025
Analytical Method: Spectrophotometry Run Batch: LB134713
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Cyanide	0	0.0	0	5	4.7	-6	10	9.3	-7

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51900 MA No.: _____ SDG No.: ME29C5
Instrument ID: KONELAB Start Date: 02/13/2025
Analytical Method: Spectrophotometry Run Batch: LB134713
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Cyanide	100	100	0	250	250	0	500	500	0