EPA SAMPLE NO.

ME29D2S

FORM 5A - IN MATRIX SPIKE SAMPLE RECOVERY

Lab Name:	Chemtech Co	onsulting Group	Contract:	EPW14030		
Lab Code:	ACE	Case No.: 51900	MA No. :		SDG No.:	ME29C5
Matrix :	Water		Analytical	Method:	CVAA	
% Solids:						
Concentrat	ion Units (μ	g/L, mg/L or mg/kg dry	weight): ug/I	L		

Analyte	Control Limit %R	Spiked Sample Result (SSR)	Q	Sample Result (SR)	Q	Spike Added (SA)	%R	Q
Mercury	75 - 125	1.3		0.2	U	1.0	126	*