

FORM 6 - IN
DUPLICATES

Lab Name: Chemtech Consulting Group Contract: EPW14030

Lab Code: ACE Case No.: 51900 MA No. : SDG No.: ME29C5

Matrix : Water Analytical Method: ICP-MS

% Solids:

Concentration Units (µg/L, mg/L or mg/kg dry weight): ug/L

Analyte	Control Limit	Sample (S)	Q	Duplicate (D)	Q	RPD	Q
Aluminum	20.0	20.0	U	19	J		
Antimony	2.0	2.0	U	0.14	J		
Arsenic	1.0	1.0	U	0.3	J		
Barium	10.0	21		24		12	
Beryllium		1.0	U	1.0	U		
Cadmium		1.0	U	1.0	U		
Calcium		240000		240000		1	
Chromium		2.0	U	2.0	U		
Cobalt	1.0	1.0	U	0.34	J		
Copper	2.0	2.0	U	2.6		200	
Iron		29	J	75	J		
Lead		0.32	J	0.6	J		
Magnesium		67000		78000		16	
Manganese		16		210		172	*
Nickel	1.0	1.0	U	2.0		200	
Potassium		2800		4000		35	*
Selenium		1.1	J	1.6	J		
Silver		1.0	U	1.0	U		
Sodium		60000		65000		8	
Thallium		1.0	U	1.0	U		
Vanadium		0.11	J	0.21	J		
Zinc	5.0	2.3	J	7.5		107	*