

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN021025

Review By	Rahul	Review On	2/11/2025 4:21:36 PM
Supervise By	Jagrut	Supervise On	2/11/2025 6:22:45 PM
SubDirectory	BN021025	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn021025
STD. NAME	STD REF.#		
Tune/Reschk	SP6717		
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731		
CCC	SP6735		
Internal Standard/PEM	SP6682,1ul/100ul sample		
ICV/I.BLK	SP6684		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN036408.D	10 Feb 2025 11:46	RC/JU	Ok
2	SSTDICC0.1	BN036409.D	10 Feb 2025 12:25	RC/JU	Ok
3	SSTDICC0.2	BN036410.D	10 Feb 2025 13:01	RC/JU	Ok
4	SSTDICCC0.4	BN036411.D	10 Feb 2025 13:36	RC/JU	Ok
5	SSTDICC0.8	BN036412.D	10 Feb 2025 14:12	RC/JU	Ok
6	SSTDICC1.6	BN036413.D	10 Feb 2025 14:48	RC/JU	Ok
7	SSTDICC3.2	BN036414.D	10 Feb 2025 15:24	RC/JU	Ok
8	SSTDICC5.0	BN036415.D	10 Feb 2025 16:00	RC/JU	Ok
9	SSTDICV0.4	BN036416.D	10 Feb 2025 16:36	RC/JU	Ok
10	PB166297BL	BN036417.D	10 Feb 2025 17:48	RC/JU	Not Ok

M : Manual Integration

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN021225

Review By	Rahul	Review On	2/13/2025 4:27:10 PM
Supervise By	Jagrut	Supervise On	2/13/2025 5:46:47 PM
SubDirectory	BN021225	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn021025
STD. NAME	STD REF.#		
Tune/Reschk	SP6717		
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731		
CCC	SP6735		
Internal Standard/PEM	SP6682,1ul/100ul sample		
ICV/I.BLK	SP6684		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN036431.D	12 Feb 2025 08:49	RC/JU	Ok
2	SSTDCCC0.4	BN036432.D	12 Feb 2025 09:34	RC/JU	Ok
3	PB166300BL	BN036433.D	12 Feb 2025 10:12	RC/JU	Ok
4	Q1168-03	BN036434.D	12 Feb 2025 10:56	RC/JU	Ok
5	Q1168-03	BN036435.D	12 Feb 2025 11:31	RC/JU	Ok,M
6	Q1168-09	BN036436.D	12 Feb 2025 12:07	RC/JU	Ok
7	Q1168-09	BN036437.D	12 Feb 2025 12:43	RC/JU	Ok
8	PB166297BL	BN036438.D	12 Feb 2025 13:19	RC/JU	Ok
9	SSTDCCC0.4	BN036439.D	12 Feb 2025 14:30	RC/JU	Ok
10	DFTPP	BN036440.D	12 Feb 2025 15:09	RC/JU	Ok
11	SSTDCCC0.4	BN036441.D	12 Feb 2025 15:48	RC/JU	Ok
12	PB166675BL	BN036442.D	12 Feb 2025 16:24	RC/JU	Ok
13	PB166609BL	BN036443.D	12 Feb 2025 17:00	RC/JU	Ok
14	Q1347-01	BN036444.D	12 Feb 2025 17:36	RC/JU	Ok
15	Q1347-03	BN036445.D	12 Feb 2025 18:12	RC/JU	Ok
16	Q1347-05	BN036446.D	12 Feb 2025 18:48	RC/JU	Ok
17	Q1325-02	BN036447.D	12 Feb 2025 19:23	RC/JU	ReRun
18	Q1325-06	BN036448.D	12 Feb 2025 19:59	RC/JU	Ok
19	Q1327-01	BN036449.D	12 Feb 2025 20:35	RC/JU	ReRun
20	Q1327-05	BN036450.D	12 Feb 2025 21:12	RC/JU	Ok,M
21	Q1327-07	BN036451.D	12 Feb 2025 21:47	RC/JU	Ok

Instrument ID: **BNA_N**

Daily Analysis Runlog For Sequence/QC Batch ID # BN021225

Review By	Rahul	Review On	2/13/2025 4:27:10 PM
Supervise By	Jagrut	Supervise On	2/13/2025 5:46:47 PM
SubDirectory	BN021225	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn021025
STD. NAME	STD REF.#		
Tune/Reschk	SP6717		
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731		
CCC	SP6735		
Internal Standard/PEM	SP6682,1ul/100ul sample		
ICV/I.BLK	SP6684		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	Q1327-09	BN036452.D	12 Feb 2025 22:23	RC/JU	Ok
23	PB166609BS	BN036453.D	12 Feb 2025 22:59	RC/JU	Ok,M
24	PB166609BSD	BN036454.D	12 Feb 2025 23:35	RC/JU	Ok,M
25	PB166675BSD	BN036455.D	13 Feb 2025 00:11	RC/JU	Ok,M
26	PB166675BS	BN036456.D	13 Feb 2025 00:47	RC/JU	Ok,M
27	SSTDCCC0.4	BN036457.D	13 Feb 2025 01:23	RC/JU	Ok,M

M : Manual Integration