

## Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1347

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN036456.D

| Lab Sample ID | Parameter   | Spike | Result | Unit | Rec | RPD | Qual | RPD  | Low | Limits | RPD |
|---------------|-------------|-------|--------|------|-----|-----|------|------|-----|--------|-----|
|               |             |       |        |      |     |     |      | Qual |     | High   |     |
| PB166675BS    | 1,4-Dioxane | 0.4   | 0.33   | ug/L | 83  |     |      |      | 70  | 130    |     |