

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1347 SAS No.: Q1347 SDG No.: Q1347

Instrument ID: MSVOA_X Calibration Date/Time: 02/12/2025 10:07

Lab File ID: VX044920.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.854	0.820	0.1	-3.98	20
Vinyl Chloride	0.827	0.797		-3.63	20
Bromomethane	0.247	0.256		3.64	20
Chloroethane	0.307	0.389		26.71	20
Trichlorofluoromethane	1.046	1.056		0.96	20
1,1,2-Trichlorotrifluoroethane	0.632	0.642		1.58	20
1,1-Dichloroethene	0.644	0.613		-4.81	20
Acetone	0.294	0.293		-0.34	20
Carbon Disulfide	1.767	1.633		-7.58	20
Methyl tert-butyl Ether	2.050	1.992		-2.83	20
Methylene Chloride	0.721	0.687		-4.72	20
trans-1,2-Dichloroethene	0.634	0.601		-5.2	20
1,1-Dichloroethane	1.233	1.202	0.1	-2.51	20
2-Butanone	0.478	0.466		-2.51	20
Carbon Tetrachloride	0.460	0.463		0.65	20
cis-1,2-Dichloroethene	0.762	0.733		-3.81	20
Chloroform	1.196	1.159		-3.09	20
1,1,1-Trichloroethane	1.014	0.988		-2.56	20
Methylcyclohexane	0.606	0.636		4.95	20
Benzene	1.465	1.458		-0.48	20
1,2-Dichloroethane	0.467	0.483		3.43	20
Trichloroethene	0.335	0.330		-1.49	20
1,2-Dichloropropane	0.364	0.366		0.55	20
Bromodichloromethane	0.489	0.505		3.27	20
4-Methyl-2-Pentanone	0.512	0.531		3.71	20
Toluene	0.872	0.887		1.72	20
t-1,3-Dichloropropene	0.495	0.514		3.84	20
cis-1,3-Dichloropropene	0.552	0.566		2.54	20
1,1,2-Trichloroethane	0.335	0.334		-0.3	20
2-Hexanone	0.369	0.382		3.52	20
Dibromochloromethane	0.359	0.367		2.23	20
Tetrachloroethene	0.315	0.319		1.27	20
Chlorobenzene	1.074	1.085	0.3	1.02	20
Ethyl Benzene	1.895	1.934		2.06	20
m/p-Xylenes	0.699	0.715		2.29	20
o-Xylene	0.701	0.711		1.43	20
Styrene	1.130	1.191		5.4	20
Bromoform	0.258	0.269	0.1	4.26	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1347 SAS No.: Q1347 SDG No.: Q1347

Instrument ID: MSVOA_X Calibration Date/Time: 02/12/2025 10:07

Lab File ID: VX044920.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	4.026	4.000		-0.65	20
1,1,2,2-Tetrachloroethane	1.383	1.289	0.3	-6.8	20
1,3-Dichlorobenzene	1.678	1.658		-1.19	20
1,4-Dichlorobenzene	1.697	1.647		-2.95	20
1,2-Dichlorobenzene	1.650	1.611		-2.36	20
1,2-Dichloroethane-d4	0.732	0.766		4.64	20
Dibromofluoromethane	0.325	0.353		8.61	20
Toluene-d8	1.229	1.313		6.84	20
4-Bromofluorobenzene	0.414	0.448		8.21	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.