

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1347 SAS No.: Q1347 SDG No.: Q1347

Instrument ID: MSVOA_X Calibration Date/Time: 02/12/2025 20:11

Lab File ID: VX044946.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.854	0.806	0.1	-5.62	50
Vinyl Chloride	0.827	0.747		-9.67	50
Bromomethane	0.247	0.253		2.43	50
Chloroethane	0.307	0.406		32.25	50
Trichlorofluoromethane	1.046	1.013		-3.15	50
1,1,2-Trichlorotrifluoroethane	0.632	0.589		-6.8	50
1,1-Dichloroethene	0.644	0.602		-6.52	50
Acetone	0.294	0.304		3.4	50
Carbon Disulfide	1.767	1.520		-13.98	50
Methyl tert-butyl Ether	2.050	1.979		-3.46	50
Methylene Chloride	0.721	0.685		-4.99	50
trans-1,2-Dichloroethene	0.634	0.607		-4.26	50
1,1-Dichloroethane	1.233	1.191	0.1	-3.41	50
2-Butanone	0.478	0.500		4.6	50
Carbon Tetrachloride	0.460	0.437		-5	50
cis-1,2-Dichloroethene	0.762	0.739		-3.02	50
Chloroform	1.196	1.171		-2.09	50
1,1,1-Trichloroethane	1.014	0.973		-4.04	50
Methylcyclohexane	0.606	0.564		-6.93	50
Benzene	1.465	1.409		-3.82	50
1,2-Dichloroethane	0.467	0.478		2.36	50
Trichloroethene	0.335	0.311		-7.16	50
1,2-Dichloropropane	0.364	0.353		-3.02	50
Bromodichloromethane	0.489	0.487		-0.41	50
4-Methyl-2-Pentanone	0.512	0.552		7.81	50
Toluene	0.872	0.851		-2.41	50
t-1,3-Dichloropropene	0.495	0.468		-5.45	50
cis-1,3-Dichloropropene	0.552	0.530		-3.99	50
1,1,2-Trichloroethane	0.335	0.336		0.3	50
2-Hexanone	0.369	0.410		11.11	50
Dibromochloromethane	0.359	0.353		-1.67	50
Tetrachloroethene	0.315	0.294		-6.67	50
Chlorobenzene	1.074	1.019	0.3	-5.12	50
Ethyl Benzene	1.895	1.798		-5.12	50
m/p-Xylenes	0.699	0.671		-4.01	50
o-Xylene	0.701	0.666		-4.99	50
Styrene	1.130	1.122		-0.71	50
Bromoform	0.258	0.248	0.1	-3.88	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1347 SAS No.: Q1347 SDG No.: Q1347

Instrument ID: MSVOA_X Calibration Date/Time: 02/12/2025 20:11

Lab File ID: VX044946.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	4.026	3.804		-5.51	50
1,1,2,2-Tetrachloroethane	1.383	1.299	0.3	-6.07	50
1,3-Dichlorobenzene	1.678	1.564		-6.79	50
1,4-Dichlorobenzene	1.697	1.566		-7.72	50
1,2-Dichlorobenzene	1.650	1.569		-4.91	50
1,2-Dichloroethane-d4	0.732	0.710		-3.01	50
Dibromofluoromethane	0.325	0.305		-6.15	50
Toluene-d8	1.229	1.123		-8.63	50
4-Bromofluorobenzene	0.414	0.400		-3.38	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.