

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1347

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VX044924.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0212WBSD01	Chloromethane	20	17.6	ug/L	88	4		50	139	20
	Vinyl chloride	20	16.9	ug/L	85	5		58	137	20
	Bromomethane	20	19.4	ug/L	97	4		53	141	20
	Chloroethane	20	22.0	ug/L	110	1		60	138	20
	Trichlorofluoromethane	20	18.6	ug/L	93	2		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	18.6	ug/L	93	2		70	136	20
	1,1-Dichloroethene	20	17.5	ug/L	88	4		71	131	20
	Acetone	100	98.2	ug/L	98	3		39	160	20
	Carbon disulfide	20	16.6	ug/L	83	2		64	133	20
	Methyl tert-butyl Ether	20	18.7	ug/L	94	1		71	124	20
	Methylene Chloride	20	18.4	ug/L	92	0		74	124	20
	trans-1,2-Dichloroethene	20	18.0	ug/L	90	1		75	124	20
	1,1-Dichloroethane	20	18.7	ug/L	94	1		77	125	20
	2-Butanone	100	99.3	ug/L	99	2		56	143	20
	Carbon Tetrachloride	20	18.7	ug/L	94	2		72	136	20
	cis-1,2-Dichloroethene	20	18.5	ug/L	93	2		78	123	20
	Chloroform	20	18.6	ug/L	93	3		79	124	20
	1,1,1-Trichloroethane	20	18.2	ug/L	91	3		74	131	20
	Methylcyclohexane	20	19.4	ug/L	97	3		72	132	20
	Benzene	20	18.9	ug/L	95	0		79	120	20
	1,2-Dichloroethane	20	20.0	ug/L	100	1		73	128	20
	Trichloroethene	20	18.7	ug/L	94	3		79	123	20
	1,2-Dichloropropane	20	19.6	ug/L	98	5		78	122	20
	Bromodichloromethane	20	19.3	ug/L	97	1		79	125	20
	4-Methyl-2-Pentanone	100	110	ug/L	110	10		67	130	20
	Toluene	20	19.4	ug/L	97	1		80	121	20
	t-1,3-Dichloropropene	20	17.8	ug/L	89	1		73	127	20
	cis-1,3-Dichloropropene	20	19.3	ug/L	97	6		75	124	20
	1,1,2-Trichloroethane	20	20.3	ug/L	102	6		80	119	20
	2-Hexanone	100	110	ug/L	110	10		57	139	20
	Dibromochloromethane	20	19.0	ug/L	95	1		74	126	20
	Tetrachloroethene	20	19.0	ug/L	95	0		74	129	20
	Chlorobenzene	20	19.0	ug/L	95	0		82	118	20
	Ethyl Benzene	20	19.0	ug/L	95	0		79	121	20
	m/p-Xylenes	40	38.9	ug/L	97	1		80	121	20
	o-Xylene	20	19.5	ug/L	98	3		78	122	20
	Styrene	20	19.7	ug/L	99	2		78	123	20
	Bromoform	20	19.4	ug/L	97	2		66	130	20
	Isopropylbenzene	20	18.7	ug/L	94	2		72	131	20
	1,1,2,2-Tetrachloroethane	20	18.2	ug/L	91	2		71	121	20
	1,3-Dichlorobenzene	20	19.3	ug/L	97	4		80	119	20
	1,4-Dichlorobenzene	20	18.5	ug/L	93	0		79	118	20
	1,2-Dichlorobenzene	20	19.1	ug/L	96	1		80	119	20