

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1347 SAS No.: Q1347 SDG NO.: Q1347
Lab File ID: VX044920.D Date Analyzed: 02/12/2025
Instrument ID: MSVOA_X Time Analyzed: 10:07
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	84033	12.018				
UPPER LIMIT	168066	12.518				
LOWER LIMIT	42016.5	11.518				
EPA SAMPLE NO.						
BP-VPB-192-EB-20250207	69850	12.02				
BP-VPB-192-TB-20250206	68278	12.02				
BP-VPB-192-GW-710-712	62921	12.02				
BP-VPB-192-GW-640-642	70282	12.02				
BP-VPB-192-GW-660-662	70815	12.02				
BP-VPB-192-GW-680-682	73243	12.02				
VX0212WBL01	76229	12.02				
VX0212WBS01	82301	12.02				
VX0212WBSD01	79611	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.