

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: Q1352

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB134675

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV pH	pH	6.98	7	100	90-110	02/12/2025
Sample ID: CCV1 pH	pH	2.01	2.00	101	90-110	02/12/2025
Sample ID: CCV2 pH	pH	12.02	12.00	100	90-110	02/12/2025

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: Q1352

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB134677

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Cyanide	mg/L	0.096	0.099	97	90-110	02/11/2025
Sample ID: CCV1 Cyanide	mg/L	0.25	0.25	100	90-110	02/11/2025
Sample ID: CCV2 Cyanide	mg/L	0.25	0.25	100	90-110	02/11/2025
Sample ID: CCV3 Cyanide	mg/L	0.25	0.25	100	90-110	02/11/2025
Sample ID: CCV4 Cyanide	mg/L	0.25	0.25	100	90-110	02/11/2025

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: Q1352

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB134677

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------