



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Weston Solutions **SDG No.:** Q1352
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1352 **SAS No.:** Q1352
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.21	0.2	105	40 - 160	CV	02/14/2025	11:10	LB134719
CRI01	Arsenic	20.3	20.0	101	40 - 160	P	02/17/2025	12:00	LB134738
	Barium	82.5	100	82	40 - 160	P	02/17/2025	12:00	LB134738
	Cadmium	5.89	6.0	98	40 - 160	P	02/17/2025	12:00	LB134738
	Chromium	9.87	10.0	99	40 - 160	P	02/17/2025	12:00	LB134738
	Lead	11.6	12.0	97	40 - 160	P	02/17/2025	12:00	LB134738
	Selenium	18.9	20.0	94	40 - 160	P	02/17/2025	12:00	LB134738
	Silver	10.3	10.0	103	40 - 160	P	02/17/2025	12:00	LB134738
CRI01	Arsenic	21.5	20.0	107	40 - 160	P	02/25/2025	16:10	lb134801
	Barium	88.7	100	89	40 - 160	P	02/25/2025	16:10	lb134801
	Cadmium	5.96	6.0	99	40 - 160	P	02/25/2025	16:10	lb134801
	Chromium	10.3	10.0	103	40 - 160	P	02/25/2025	16:10	lb134801
	Lead	11.3	12.0	94	40 - 160	P	02/25/2025	16:10	lb134801
	Selenium	18.7	20.0	94	40 - 160	P	02/25/2025	16:10	lb134801
	Silver	10.5	10.0	105	40 - 160	P	02/25/2025	16:10	lb134801