



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>Weston Solutions</u>				SDG No.:	<u>Q1352</u>				
Contract:	<u>WEST04</u>	Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1352</u>	SAS No.:	<u>Q1352</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB05	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/14/2025	11:03	LB134719

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Weston Solutions

SDG No.: Q1352

Contract: WEST04

Lab Code: CHEM

Case No.: Q1352

SAS No.: Q1352

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB72	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/14/2025	11:08	LB134719
CCB73	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/14/2025	11:38	LB134719
CCB74	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/14/2025	12:08	LB134719
CCB75	Mercury	0.20	+/-0.20	U	0.16	0.20	CV	02/14/2025	12:38	LB134719

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>Q1352</u>								
Contract: <u>WEST04</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q1352</u>			SAS No.: <u>Q1352</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	02/17/2025	11:56	LB134738
	Barium	100	+/-100	U	25.0	100	P	02/17/2025	11:56	LB134738
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	02/17/2025	11:56	LB134738
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	02/17/2025	11:56	LB134738
	Lead	12.0	+/-12.0	U	9.60	12.0	P	02/17/2025	11:56	LB134738
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	02/17/2025	11:56	LB134738
	Silver	10.0	+/-10.0	U	5.00	10.0	P	02/17/2025	11:56	LB134738

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>Q1352</u>								
Contract: <u>WEST04</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1352</u>	SAS No.: <u>Q1352</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	02/17/2025	12:17	LB134738
	Barium	100	+/-100	U	25.0	100	P	02/17/2025	12:17	LB134738
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	02/17/2025	12:17	LB134738
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	02/17/2025	12:17	LB134738
	Lead	12.0	+/-12.0	U	9.60	12.0	P	02/17/2025	12:17	LB134738
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	02/17/2025	12:17	LB134738
	Silver	10.0	+/-10.0	U	5.00	10.0	P	02/17/2025	12:17	LB134738
CCB02	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	02/17/2025	13:10	LB134738
	Barium	100	+/-100	U	25.0	100	P	02/17/2025	13:10	LB134738
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	02/17/2025	13:10	LB134738
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	02/17/2025	13:10	LB134738
	Lead	12.0	+/-12.0	U	9.60	12.0	P	02/17/2025	13:10	LB134738
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	02/17/2025	13:10	LB134738
	Silver	10.0	+/-10.0	U	5.00	10.0	P	02/17/2025	13:10	LB134738
CCB03	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	02/17/2025	14:06	LB134738
	Barium	100	+/-100	U	25.0	100	P	02/17/2025	14:06	LB134738
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	02/17/2025	14:06	LB134738
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	02/17/2025	14:06	LB134738
	Lead	12.0	+/-12.0	U	9.60	12.0	P	02/17/2025	14:06	LB134738
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	02/17/2025	14:06	LB134738
	Silver	10.0	+/-10.0	U	5.00	10.0	P	02/17/2025	14:06	LB134738
CCB04	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	02/17/2025	14:58	LB134738
	Barium	100	+/-100	U	25.0	100	P	02/17/2025	14:58	LB134738
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	02/17/2025	14:58	LB134738
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	02/17/2025	14:58	LB134738
	Lead	12.0	+/-12.0	U	9.60	12.0	P	02/17/2025	14:58	LB134738
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	02/17/2025	14:58	LB134738
	Silver	10.0	+/-10.0	U	5.00	10.0	P	02/17/2025	14:58	LB134738
CCB05	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	02/17/2025	15:15	LB134738
	Barium	100	+/-100	U	25.0	100	P	02/17/2025	15:15	LB134738
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	02/17/2025	15:15	LB134738
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	02/17/2025	15:15	LB134738
	Lead	12.0	+/-12.0	U	9.60	12.0	P	02/17/2025	15:15	LB134738
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	02/17/2025	15:15	LB134738
	Silver	10.0	+/-10.0	U	5.00	10.0	P	02/17/2025	15:15	LB134738

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>Q1352</u>								
Contract: <u>WEST04</u>		Lab Code: <u>CHEM</u>		Case No.: <u>Q1352</u>			SAS No.: <u>Q1352</u>			
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	02/25/2025	16:06	lb134801
	Barium	100	+/-100	U	25.0	100	P	02/25/2025	16:06	lb134801
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	02/25/2025	16:06	lb134801
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	02/25/2025	16:06	lb134801
	Lead	12.0	+/-12.0	U	9.60	12.0	P	02/25/2025	16:06	lb134801
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	02/25/2025	16:06	lb134801
	Silver	10.0	+/-10.0	U	5.00	10.0	P	02/25/2025	16:06	lb134801

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Weston Solutions</u>		SDG No.: <u>Q1352</u>								
Contract: <u>WEST04</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1352</u>	SAS No.: <u>Q1352</u>							
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	LOD	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	02/25/2025	16:27	lb134801
	Barium	100	+/-100	U	25.0	100	P	02/25/2025	16:27	lb134801
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	02/25/2025	16:27	lb134801
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	02/25/2025	16:27	lb134801
	Lead	12.0	+/-12.0	U	9.60	12.0	P	02/25/2025	16:27	lb134801
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	02/25/2025	16:27	lb134801
	Silver	10.0	+/-10.0	U	5.00	10.0	P	02/25/2025	16:27	lb134801
CCB02	Arsenic	20.0	+/-20.0	U	16.0	20.0	P	02/25/2025	16:58	lb134801
	Barium	100	+/-100	U	25.0	100	P	02/25/2025	16:58	lb134801
	Cadmium	6.00	+/-6.00	U	1.50	6.00	P	02/25/2025	16:58	lb134801
	Chromium	10.0	+/-10.0	U	5.00	10.0	P	02/25/2025	16:58	lb134801
	Lead	12.0	+/-12.0	U	9.60	12.0	P	02/25/2025	16:58	lb134801
	Selenium	20.0	+/-20.0	U	16.0	20.0	P	02/25/2025	16:58	lb134801
	Silver	10.0	+/-10.0	U	5.00	10.0	P	02/25/2025	16:58	lb134801