

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions **SDG No.:** Q1352
Contract: WEST04 **Lab Code:** CHEM **Case No.:** Q1352 **SAS No.:** Q1352

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166715BS							
Arsenic	ug/L	4000	3980		100	87 - 113	P
Barium	ug/L	1000	1010		101	88 - 113	P
Cadmium	ug/L	1000	953		95	88 - 113	P
Chromium	ug/L	2000	2030		102	90 - 113	P
Lead	ug/L	5000	4820		96	86 - 113	P
Selenium	ug/L	10000	9580		96	83 - 114	P
Silver	ug/L	380	379		100	84 - 115	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: Weston Solutions

SDG No.: Q1352

Contract: WEST04

Lab Code: CHEM

Case No.: Q1352

SAS No.: Q1352

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166720BS Mercury	ug/L	4.0	3.60		90	82 - 119	CV