

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL01 Date Analyzed: 02/12/2025

Lab Sample No.: AR1660CCC500 Data File : PP069670.D Time Analyzed: 09:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.689	5.592	5.792	514.840	500.000	3.0
Aroclor-1016-2	5.710	5.614	5.814	524.890	500.000	5.0
Aroclor-1016-3	5.773	5.676	5.876	510.710	500.000	2.1
Aroclor-1016-4	5.871	5.774	5.974	521.610	500.000	4.3
Aroclor-1016-5	6.164	6.067	6.267	512.630	500.000	2.5
Aroclor-1260-1	7.282	7.187	7.387	453.200	500.000	-9.4
Aroclor-1260-2	7.536	7.441	7.641	449.310	500.000	-10.1
Aroclor-1260-3	7.895	7.800	8.000	454.030	500.000	-9.2
Aroclor-1260-4	8.119	8.025	8.225	423.630	500.000	-15.3
Aroclor-1260-5	8.440	8.347	8.547	443.470	500.000	-11.3
Decachlorobiphenyl	10.269	10.179	10.379	48.630	50.000	-2.7
Tetrachloro-m-xylene	4.536	4.438	4.638	54.300	50.000	8.6

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL01 Date Analyzed: 02/12/2025

Lab Sample No.: AR1660CCC500 Data File : PP069670.D Time Analyzed: 09:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.929	4.834	5.034	517.040	500.000	3.4
Aroclor-1016-2	4.948	4.853	5.053	523.770	500.000	4.8
Aroclor-1016-3	5.126	5.030	5.230	527.310	500.000	5.5
Aroclor-1016-4	5.167	5.072	5.272	515.870	500.000	3.2
Aroclor-1016-5	5.382	5.288	5.488	527.210	500.000	5.4
Aroclor-1260-1	6.421	6.326	6.526	523.760	500.000	4.8
Aroclor-1260-2	6.608	6.515	6.715	516.240	500.000	3.2
Aroclor-1260-3	6.763	6.670	6.870	454.590	500.000	-9.1
Aroclor-1260-4	7.235	7.142	7.342	495.190	500.000	-1.0
Aroclor-1260-5	7.476	7.384	7.584	496.840	500.000	-0.6
Decachlorobiphenyl	8.900	8.811	9.011	46.250	50.000	-7.5
Tetrachloro-m-xylene	3.839	3.740	3.940	53.530	50.000	7.1