

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 02/12/2025

Lab Sample No.: AR1660CCC500 Data File : PP069684.D Time Analyzed: 15:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.686	5.592	5.792	516.380	500.000	3.3
Aroclor-1016-2	5.707	5.614	5.814	511.390	500.000	2.3
Aroclor-1016-3	5.769	5.676	5.876	504.930	500.000	1.0
Aroclor-1016-4	5.867	5.774	5.974	513.380	500.000	2.7
Aroclor-1016-5	6.160	6.067	6.267	505.230	500.000	1.0
Aroclor-1260-1	7.278	7.187	7.387	477.620	500.000	-4.5
Aroclor-1260-2	7.532	7.441	7.641	450.670	500.000	-9.9
Aroclor-1260-3	7.891	7.800	8.000	450.580	500.000	-9.9
Aroclor-1260-4	8.115	8.025	8.225	428.740	500.000	-14.3
Aroclor-1260-5	8.435	8.347	8.547	446.270	500.000	-10.7
Decachlorobiphenyl	10.262	10.179	10.379	49.000	50.000	-2.0
Tetrachloro-m-xylene	4.533	4.438	4.638	53.170	50.000	6.3

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 02/12/2025

Lab Sample No.: AR1660CCC500 Data File : PP069684.D Time Analyzed: 15:59

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.927	4.834	5.034	501.590	500.000	0.3
Aroclor-1016-2	4.946	4.853	5.053	508.320	500.000	1.7
Aroclor-1016-3	5.123	5.030	5.230	509.900	500.000	2.0
Aroclor-1016-4	5.164	5.072	5.272	504.660	500.000	0.9
Aroclor-1016-5	5.380	5.288	5.488	503.820	500.000	0.8
Aroclor-1260-1	6.418	6.326	6.526	515.090	500.000	3.0
Aroclor-1260-2	6.606	6.515	6.715	508.090	500.000	1.6
Aroclor-1260-3	6.759	6.670	6.870	469.830	500.000	-6.0
Aroclor-1260-4	7.232	7.142	7.342	489.630	500.000	-2.1
Aroclor-1260-5	7.471	7.384	7.584	492.260	500.000	-1.5
Decachlorobiphenyl	8.895	8.811	9.011	45.760	50.000	-8.5
Tetrachloro-m-xylene	3.837	3.740	3.940	51.700	50.000	3.4