

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL03 Date Analyzed: 02/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP069716.D Time Analyzed: 13:47

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.687	5.592	5.792	552.380	500.000	10.5
Aroclor-1016-2	5.708	5.614	5.814	557.950	500.000	11.6
Aroclor-1016-3	5.771	5.676	5.876	539.580	500.000	7.9
Aroclor-1016-4	5.868	5.774	5.974	562.000	500.000	12.4
Aroclor-1016-5	6.161	6.067	6.267	534.400	500.000	6.9
Aroclor-1260-1	7.281	7.187	7.387	538.070	500.000	7.6
Aroclor-1260-2	7.535	7.441	7.641	519.130	500.000	3.8
Aroclor-1260-3	7.894	7.800	8.000	526.080	500.000	5.2
Aroclor-1260-4	8.119	8.025	8.225	503.560	500.000	0.7
Aroclor-1260-5	8.439	8.347	8.547	508.820	500.000	1.8
Decachlorobiphenyl	10.269	10.179	10.379	53.670	50.000	7.3
Tetrachloro-m-xylene	4.534	4.438	4.638	57.480	50.000	15.0

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL03 Date Analyzed: 02/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP069716.D Time Analyzed: 13:47

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.928	4.834	5.034	541.970	500.000	8.4
Aroclor-1016-2	4.947	4.853	5.053	549.830	500.000	10.0
Aroclor-1016-3	5.125	5.030	5.230	545.340	500.000	9.1
Aroclor-1016-4	5.166	5.072	5.272	526.800	500.000	5.4
Aroclor-1016-5	5.382	5.288	5.488	549.580	500.000	9.9
Aroclor-1260-1	6.421	6.326	6.526	569.870	500.000	14.0
Aroclor-1260-2	6.608	6.515	6.715	563.290	500.000	12.7
Aroclor-1260-3	6.762	6.670	6.870	496.210	500.000	-0.8
Aroclor-1260-4	7.235	7.142	7.342	553.350	500.000	10.7
Aroclor-1260-5	7.476	7.384	7.584	565.920	500.000	13.2
Decachlorobiphenyl	8.899	8.811	9.011	49.320	50.000	-1.4
Tetrachloro-m-xylene	3.837	3.740	3.940	56.590	50.000	13.2