

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL04 Date Analyzed: 02/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP069728.D Time Analyzed: 17:13

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.689	5.592	5.792	551.770	500.000	10.4
Aroclor-1016-2	5.710	5.614	5.814	558.930	500.000	11.8
Aroclor-1016-3	5.773	5.676	5.876	545.320	500.000	9.1
Aroclor-1016-4	5.871	5.774	5.974	557.170	500.000	11.4
Aroclor-1016-5	6.164	6.067	6.267	551.370	500.000	10.3
Aroclor-1260-1	7.283	7.187	7.387	536.480	500.000	7.3
Aroclor-1260-2	7.536	7.441	7.641	522.170	500.000	4.4
Aroclor-1260-3	7.895	7.800	8.000	527.220	500.000	5.4
Aroclor-1260-4	8.120	8.025	8.225	501.210	500.000	0.2
Aroclor-1260-5	8.440	8.347	8.547	510.440	500.000	2.1
Decachlorobiphenyl	10.269	10.179	10.379	53.970	50.000	7.9
Tetrachloro-m-xylene	4.536	4.438	4.638	57.110	50.000	14.2

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL04 Date Analyzed: 02/13/2025

Lab Sample No.: AR1660CCC500 Data File : PP069728.D Time Analyzed: 17:13

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.928	4.834	5.034	535.460	500.000	7.1
Aroclor-1016-2	4.947	4.853	5.053	541.410	500.000	8.3
Aroclor-1016-3	5.125	5.030	5.230	538.950	500.000	7.8
Aroclor-1016-4	5.167	5.072	5.272	512.100	500.000	2.4
Aroclor-1016-5	5.382	5.288	5.488	555.040	500.000	11.0
Aroclor-1260-1	6.420	6.326	6.526	566.660	500.000	13.3
Aroclor-1260-2	6.608	6.515	6.715	562.260	500.000	12.5
Aroclor-1260-3	6.762	6.670	6.870	496.710	500.000	-0.7
Aroclor-1260-4	7.235	7.142	7.342	558.960	500.000	11.8
Aroclor-1260-5	7.475	7.384	7.584	562.320	500.000	12.5
Decachlorobiphenyl	8.899	8.811	9.011	50.990	50.000	2.0
Tetrachloro-m-xylene	3.838	3.740	3.940	55.440	50.000	10.9