

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP012825

Review By	yogesh	Review On	1/29/2025 8:01:22 AM
Supervise By	Ankita	Supervise On	1/29/2025 12:56:57 PM
SubDirectory	PP012825	HP Acquire Method	HP Processing Method PP012825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP069263.D	28 Jan 2025 09:05		YPIAJ	Ok
2	I.BLK	I.BLK	PP069264.D	28 Jan 2025 09:21		YPIAJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PP069265.D	28 Jan 2025 09:37		YPIAJ	Ok,M
4	AR1660ICC750	AR1660ICC750	PP069266.D	28 Jan 2025 09:54		YPIAJ	Ok
5	AR1660ICC500	AR1660ICC500	PP069267.D	28 Jan 2025 10:10		YPIAJ	Ok
6	AR1660ICC250	AR1660ICC250	PP069268.D	28 Jan 2025 10:26		YPIAJ	Ok
7	AR1660ICC050	AR1660ICC050	PP069269.D	28 Jan 2025 11:15		YPIAJ	Ok
8	AR1221ICC500	AR1221ICC500	PP069270.D	28 Jan 2025 11:32		YPIAJ	Ok
9	AR1232ICC500	AR1232ICC500	PP069271.D	28 Jan 2025 11:48		YPIAJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PP069272.D	28 Jan 2025 12:04		YPIAJ	Ok
11	AR1242ICC750	AR1242ICC750	PP069273.D	28 Jan 2025 12:21		YPIAJ	Ok
12	AR1242ICC500	AR1242ICC500	PP069274.D	28 Jan 2025 12:37		YPIAJ	Ok
13	AR1242ICC250	AR1242ICC250	PP069275.D	28 Jan 2025 12:53		YPIAJ	Ok
14	AR1242ICC050	AR1242ICC050	PP069276.D	28 Jan 2025 13:09		YPIAJ	Ok
15	AR1248ICC1000	AR1248ICC1000	PP069277.D	28 Jan 2025 13:26		YPIAJ	Ok
16	AR1248ICC750	AR1248ICC750	PP069278.D	28 Jan 2025 13:42		YPIAJ	Ok
17	AR1248ICC500	AR1248ICC500	PP069279.D	28 Jan 2025 13:58		YPIAJ	Ok
18	AR1248ICC250	AR1248ICC250	PP069280.D	28 Jan 2025 14:15		YPIAJ	Ok,M

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP012825

Review By	yogesh	Review On	1/29/2025 8:01:22 AM
Supervise By	Ankita	Supervise On	1/29/2025 12:56:57 PM
SubDirectory	PP012825	HP Acquire Method	HP Processing Method PP012825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PP069281.D	28 Jan 2025 14:31		YPIAJ	Ok
20	AR1254ICC1000	AR1254ICC1000	PP069282.D	28 Jan 2025 14:47		YPIAJ	Ok
21	AR1254ICC750	AR1254ICC750	PP069283.D	28 Jan 2025 15:04		YPIAJ	Ok
22	AR1254ICC500	AR1254ICC500	PP069284.D	28 Jan 2025 15:20		YPIAJ	Ok
23	AR1254ICC250	AR1254ICC250	PP069285.D	28 Jan 2025 15:36		YPIAJ	Ok
24	AR1254ICC050	AR1254ICC050	PP069286.D	28 Jan 2025 15:53		YPIAJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PP069287.D	28 Jan 2025 16:09		YPIAJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PP069288.D	28 Jan 2025 16:25		YPIAJ	Ok
27	AR1268ICC750	AR1268ICC750	PP069289.D	28 Jan 2025 16:42		YPIAJ	Ok
28	AR1268ICC500	AR1268ICC500	PP069290.D	28 Jan 2025 16:58		YPIAJ	Ok
29	AR1268ICC250	AR1268ICC250	PP069291.D	28 Jan 2025 17:14		YPIAJ	Ok
30	AR1268ICC050	AR1268ICC050	PP069292.D	28 Jan 2025 17:30		YPIAJ	Ok,M
31	PP012825ICV500	ICVPP012825	PP069293.D	28 Jan 2025 17:47		YPIAJ	Ok
32	AR1242ICV500	ICVPP012825AR1242	PP069294.D	28 Jan 2025 18:03		YPIAJ	Ok
33	AR1248ICV500	ICVPP012825AR1248	PP069295.D	28 Jan 2025 18:19		YPIAJ	Ok
34	AR1254ICV500	ICVPP012825AR1254	PP069296.D	28 Jan 2025 18:36		YPIAJ	Ok
35	AR1268ICV500	ICVPP012825AR1268	PP069297.D	28 Jan 2025 18:52		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP021225

Review By	yogesh	Review On	2/13/2025 7:50:13 AM
Supervise By	Ankita	Supervise On	2/13/2025 9:53:18 AM
SubDirectory	PP021225	HP Acquire Method	HP Processing Method PP012825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP069669.D	12 Feb 2025 09:22		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP069670.D	12 Feb 2025 09:38		YPIAJ	Ok,M
3	AR1242CCC500	AR1242CCC500	PP069671.D	12 Feb 2025 09:54		YPIAJ	Ok
4	AR1248CCC500	AR1248CCC500	PP069672.D	12 Feb 2025 10:10		YPIAJ	Ok
5	AR1254CCC500	AR1254CCC500	PP069673.D	12 Feb 2025 10:27		YPIAJ	Ok
6	I.BLK	I.BLK	PP069674.D	12 Feb 2025 10:43		YPIAJ	Ok
7	PB166696BL	PB166696BL	PP069675.D	12 Feb 2025 13:12	TCMX high both column	YPIAJ	Not Ok
8	PB166696BS	PB166696BS	PP069676.D	12 Feb 2025 13:28	TCMX high 1st column	YPIAJ	Not Ok
9	Q1352-01	TAP-IDW-SOIL-021025	PP069677.D	12 Feb 2025 13:45		YPIAJ	Ok
10	Q1353-01	346	PP069678.D	12 Feb 2025 14:01		YPIAJ	Ok,M
11	Q1354-01	NB-08-021125	PP069679.D	12 Feb 2025 14:17	AR1254 Hit	YPIAJ	Ok,M
12	Q1356-01	CARBON-SOLID	PP069680.D	12 Feb 2025 14:33	DCB low in both column	YPIAJ	ReRun
13	Q1356-03	SOIL-PILE	PP069681.D	12 Feb 2025 14:50	AR1248 Hit,	YPIAJ	Ok,M
14	Q1356-03MS	SOIL-PILEMS	PP069682.D	12 Feb 2025 15:06	Recovery fail for AR1016 & AR1260	YPIAJ	Ok,M
15	Q1356-03MSD	SOIL-PILEMSD	PP069683.D	12 Feb 2025 15:22	Recovery fail for AR1016 & AR1260	YPIAJ	Ok,M
16	AR1660CCC500	AR1660CCC500	PP069684.D	12 Feb 2025 15:59		YPIAJ	Ok,M
17	AR1242CCC500	AR1242CCC500	PP069685.D	12 Feb 2025 16:16		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP021225

Review By	yogesh	Review On	2/13/2025 7:50:13 AM
Supervise By	Ankita	Supervise On	2/13/2025 9:53:18 AM
SubDirectory	PP021225	HP Acquire Method	HP Processing Method PP012825
STD. NAME	STD REF.#		
Tune/Reschk			
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775		
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773		
Internal Standard/PEM			
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

18	AR1248CCC500	AR1248CCC500	PP069686.D	12 Feb 2025 16:32		YPIAJ	Ok
19	AR1254CCC500	AR1254CCC500	PP069687.D	12 Feb 2025 16:48		YPIAJ	Ok
20	I.BLK	I.BLK	PP069688.D	12 Feb 2025 17:05		YPIAJ	Ok
21	PB166695BL	PB166695BL	PP069689.D	12 Feb 2025 17:21		YPIAJ	Ok
22	PB166695BS	PB166695BS	PP069690.D	12 Feb 2025 17:37		YPIAJ	Ok,M
23	PB166695BSD	PB166695BSD	PP069691.D	12 Feb 2025 17:53		YPIAJ	Ok,M
24	Q1348-01	TWP-1-WC	PP069692.D	12 Feb 2025 18:10		YPIAJ	Ok
25	Q1356-04	CARBON-WATER	PP069693.D	12 Feb 2025 18:26		YPIAJ	Ok
26	Q1356-05	CARBON-FB	PP069694.D	12 Feb 2025 18:42		YPIAJ	Ok
27	Q1356-06	WATER-A	PP069695.D	12 Feb 2025 18:58		YPIAJ	Ok
28	Q1356-07	WATER-B	PP069696.D	12 Feb 2025 19:15		YPIAJ	Ok
29	Q1356-08	WATER-FB	PP069697.D	12 Feb 2025 19:31		YPIAJ	Ok
30	Q1356-09	SOIL-FB	PP069698.D	12 Feb 2025 19:47		YPIAJ	Ok
31	AR1660CCC500	AR1660CCC500	PP069699.D	12 Feb 2025 20:25		YPIAJ	Ok,M
32	AR1242CCC500	AR1242CCC500	PP069700.D	12 Feb 2025 20:41		YPIAJ	Ok
33	AR1248CCC500	AR1248CCC500	PP069701.D	12 Feb 2025 20:57		YPIAJ	Ok
34	AR1254CCC500	AR1254CCC500	PP069702.D	12 Feb 2025 21:14		YPIAJ	Ok
35	I.BLK	I.BLK	PP069703.D	12 Feb 2025 21:30		YPIAJ	Ok

M : Manual Integration

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP021325

Review By	yogesh	Review On	2/14/2025 7:55:07 AM
Supervise By	Ankita	Supervise On	2/14/2025 9:38:48 AM
SubDirectory	PP021325	HP Acquire Method	HP Processing Method PP012825

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773
Internal Standard/PEM	
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	Sampleld	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PP069704.D	13 Feb 2025 09:00		YPIAJ	Ok
2	AR1660CCC500	AR1660CCC500	PP069705.D	13 Feb 2025 09:16		YPIAJ	Ok
3	AR1242CCC500	AR1242CCC500	PP069706.D	13 Feb 2025 09:32		YPIAJ	Ok
4	AR1248CCC500	AR1248CCC500	PP069707.D	13 Feb 2025 09:48		YPIAJ	Ok
5	AR1254CCC500	AR1254CCC500	PP069708.D	13 Feb 2025 10:05		YPIAJ	Ok
6	I.BLK	I.BLK	PP069709.D	13 Feb 2025 10:21		YPIAJ	Ok
7	Q1356-01RE	CARBON-SOLIDRE	PP069710.D	13 Feb 2025 11:29	DCB low in both column	YPIAJ	Confirms
8	PB166708BL	PB166708BL	PP069711.D	13 Feb 2025 12:15		YPIAJ	Ok
9	PB166708BS	PB166708BS	PP069712.D	13 Feb 2025 12:32		YPIAJ	Ok,M
10	Q1358-05	5016	PP069713.D	13 Feb 2025 12:48		YPIAJ	Ok,M
11	Q1358-05MS	5016MS	PP069714.D	13 Feb 2025 13:04		YPIAJ	Ok,M
12	Q1358-05MSD	5016MSD	PP069715.D	13 Feb 2025 13:20		YPIAJ	Ok,M
13	AR1660CCC500	AR1660CCC500	PP069716.D	13 Feb 2025 13:47		YPIAJ	Ok
14	AR1242CCC500	AR1242CCC500	PP069717.D	13 Feb 2025 14:04		YPIAJ	Ok
15	AR1248CCC500	AR1248CCC500	PP069718.D	13 Feb 2025 14:20		YPIAJ	Ok
16	AR1254CCC500	AR1254CCC500	PP069719.D	13 Feb 2025 14:36		YPIAJ	Ok
17	I.BLK	I.BLK	PP069720.D	13 Feb 2025 14:53		YPIAJ	Ok
18	PB166709BL	PB166709BL	PP069721.D	13 Feb 2025 15:09		YPIAJ	Ok

Instrument ID: ECD_P

Daily Analysis Runlog For Sequence/QC Batch ID # PP021325

Review By	yogesh	Review On	2/14/2025 7:55:07 AM		
Supervise By	Ankita	Supervise On	2/14/2025 9:38:48 AM		
SubDirectory	PP021325	HP Acquire Method	HP Processing Method	PP012825	
STD. NAME	STD REF.#				
Tune/Reschk					
Initial Calibration Stds	PP23735,PP23736,PP23737,PP23738,PP23739,PP23740,PP23741,PP23742,PP23743,PP23744,PP23745,PP23747,PP23748,PP23749,PP23750,P P23751,PP23752,PP23753,PP23754,PP23755,PP23756,PP23757,PP23758,PP23759,PP23760,PP23761,PP23762,PP23763,PP23764,PP23765,PP 23766,PP23767,PP23768,PP23769,PP23770,PP23771,PP23772,PP23773,PP23774,PP23775				
CCC	PP23737,PP23742,PP23749,PP23754,PP23758,PP23763,PP23768,PP23773				
Internal Standard/PEM					
ICV/I.BLK	PP23778,PP23780,PP23783,PP23784,PP23786,PP23788,PP23790,PP23947				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	PB166709BS	PB166709BS	PP069722.D	13 Feb 2025 15:25		YPIAJ	Ok,M
20	PB166709BSD	PB166709BSD	PP069723.D	13 Feb 2025 15:41		YPIAJ	Ok,M
21	Q1357-03	NWB-2135-36	PP069724.D	13 Feb 2025 15:58		YPIAJ	Ok
22	Q1358-04	207-COMP	PP069725.D	13 Feb 2025 16:14		YPIAJ	Ok
23	PB166696BL	PB166696BL	PP069726.D	13 Feb 2025 16:30		YPIAJ	Ok
24	PB166696BS	PB166696BS	PP069727.D	13 Feb 2025 16:47		YPIAJ	Ok
25	AR1660CCC500	AR1660CCC500	PP069728.D	13 Feb 2025 17:13		YPIAJ	Ok
26	AR1242CCC500	AR1242CCC500	PP069729.D	13 Feb 2025 17:30		YPIAJ	Ok
27	AR1248CCC500	AR1248CCC500	PP069730.D	13 Feb 2025 17:46		YPIAJ	Ok
28	AR1254CCC500	AR1254CCC500	PP069731.D	13 Feb 2025 18:02		YPIAJ	Ok
29	I.BLK	I.BLK	PP069732.D	13 Feb 2025 18:19		YPIAJ	Ok

M : Manual Integration