



Contract:	WEST04						
Lab Code:	CHEM	Case No.:	Q1352	SAS No.:	Q1352	SDG NO.:	Q1352
Instrument ID:	ECD_S	Calibration Date(s):		02/11/2025	02/11/2025		
		Calibration Times:		17:56	19:32		

LAB FILE ID:		RT 200 =	<u>PS029118.D</u>	RT 500 =	<u>PS029119.D</u>
RT 750 =	<u>PS029120.D</u>	RT 1000 =	<u>PS029121.D</u>	RT 1500 =	<u>PS029122.D</u>

[illegible]

RETENTION TIMES OF INITIAL CALIBRATION

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

Instrument ID: ECD_S Calibration Date(s): 02/11/2025 02/11/2025

Calibration Times: 17:56 19:32

GC Column: RTX-CLP2 ID: 0.32 (mm)

LAB FILE ID:	RT 200 = <u>PS029118.D</u>	RT 500 = <u>PS029119.D</u>
	RT 750 = <u>PS029120.D</u>	RT 1000 = <u>PS029121.D</u>
		RT 1500 = <u>PS029122.D</u>

COMPOUND	RT 200	RT 500	RT 750	RT 1000	RT 1500	MEAN RT	RT WINDOW FROM TO	
2,4,5-TP(Silvex)	9.80	9.80	9.80	9.80	9.80	9.80	9.70	9.90
2,4-D	8.90	8.90	8.90	8.90	8.90	8.90	8.80	9.00
2,4-DCAA	7.67	7.67	7.66	7.66	7.66	7.66	7.56	7.76