

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

Instrument ID: ECD_S Calibration Date(s): 02/11/2025 02/11/2025
Calibration Times: 17:56 19:32

GC Column: RTX-CLP ID: 0.32 (mm)

LAB FILE ID:		CF 200 =	<u>PS029118.D</u>	CF 500 =	<u>PS029119.D</u>		
CF 750 =		<u>PS029120.D</u>	CF 1000 =	<u>PS029121.D</u>	CF 1500 =	<u>PS029122.D</u>	
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	23549600000	19203700000	18837600000	18189200000	17171500000	19390300000	13
2,4-D	4145320000	3324850000	3237670000	3127190000	2985430000	3364090000	14
2,4-DCAA	3443860000	2743930000	2645570000	2552190000	2433720000	2763860000	14

CALIBRATION FACTOR OF INITIAL CALIBRATION

Contract: WEST04
Lab Code: CHEM **Case No.:** Q1352 **SAS No.:** Q1352 **SDG NO.:** Q1352
Instrument ID: ECD_S **Calibration Date(s):** 02/11/2025 02/11/2025
Calibration Times: 17:56 19:32

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

LAB FILE ID:		CF 200 = <u>PS029118.D</u>	CF 500 = <u>PS029119.D</u>				
CF 750 = <u>PS029120.D</u>		CF 1000 = <u>PS029121.D</u>	CF 1500 = <u>PS029122.D</u>				
COMPOUND	CF 200	CF 500	CF 750	CF 1000	CF 1500	CF	% RSD
2,4,5-TP(Silvex)	11056300000	9541830000	9635560000	9448160000	9159180000	9768210000	8
2,4-D	1655040000	1386360000	1398110000	1372120000	1351500000	1432630000	9
2,4-DCAA	1265650000	1023410000	1020630000	1000720000	982068000	1058500000	11