

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM **Case No.:** Q1352 **SAS No.:** Q1352 **SDG NO.:** Q1352

Continuing Calib Date: 02/13/2025 **Initial Calibration Date(s):** 02/11/2025 02/11/2025

Continuing Calib Time: 21:48 **Initial Calibration Time(s):** 17:56 19:32

GC Column: RTX-CLP **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.19	7.19	7.09	7.29	0.00
2,4-D	8.30	8.30	8.20	8.40	0.00
2,4,5-TP(Silvex)	9.17	9.17	9.07	9.27	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM **Case No.:** Q1352 **SAS No.:** Q1352 **SDG NO.:** Q1352

Continuing Calib Date: 02/13/2025 **Initial Calibration Date(s):** 02/11/2025 02/11/2025

Continuing Calib Time: 21:48 **Initial Calibration Time(s):** 17:56 19:32

GC Column: RTX-CLP2 **ID:** 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
2,4-DCAA	7.66	7.66	7.56	7.76	0.00
2,4-D	8.89	8.90	8.80	9.00	0.01
2,4,5-TP(Silvex)	9.79	9.80	9.70	9.90	0.01