

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/11/2025 02/11/2025

Client Sample No.: CCAL01 Date Analyzed: 02/13/2025

Lab Sample No.: HSTDCCC750 Data File : PS029160.D Time Analyzed: 18:36

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
2,4,5-TP(Silvex)	9.175	9.073	9.273	729.100	712.500	2.3
2,4-D	8.305	8.202	8.402	711.710	705.000	1.0
2,4-DCAA	7.189	7.087	7.287	750.850	750.000	0.1

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/11/2025 02/11/2025

Client Sample No.: CCAL01 Date Analyzed: 02/13/2025

Lab Sample No.: HSTDCCC750 Data File : PS029160.D Time Analyzed: 18:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.793	9.696	9.896	718.930	712.500	0.9
2,4-D	8.895	8.797	8.997	694.290	705.000	-1.5
2,4-DCAA	7.663	7.564	7.764	727.200	750.000	-3.0