

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/11/2025 02/11/2025

Client Sample No.: CCAL02 Date Analyzed: 02/13/2025

Lab Sample No.: HSTDCCC750 Data File : PS029167.D Time Analyzed: 21:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.174	9.073	9.273	712.400	712.500	0.0
2,4-D	8.304	8.202	8.402	690.020	705.000	-2.1
2,4-DCAA	7.188	7.087	7.287	733.660	750.000	-2.2

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM **Case No.:** Q1352 **SAS No.:** Q1352 **SDG NO.:** Q1352

GC Column: RTX-CLP2 **ID:** 0.32 (mm) **Initi. Calib. Date(s):** 02/11/2025 02/11/2025

Client Sample No.: CCAL02 **Date Analyzed:** 02/13/2025

Lab Sample No.: HSTDCCC750 **Data File :** PS029167.D **Time Analyzed:** 21:48

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.793	9.696	9.896	706.270	712.500	-0.9
2,4-D	8.894	8.797	8.997	677.350	705.000	-3.9
2,4-DCAA	7.662	7.564	7.764	712.360	750.000	-5.0