

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: RTX-CLP ID: 0.32 (mm) Initi. Calib. Date(s): 02/11/2025 02/11/2025

Client Sample No.: CCAL03 Date Analyzed: 02/14/2025

Lab Sample No.: HSTDCCC750 Data File : PS029179.D Time Analyzed: 03:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.173	9.073	9.273	706.420	712.500	-0.9
2,4-D	8.303	8.202	8.402	683.610	705.000	-3.0
2,4-DCAA	7.187	7.087	7.287	723.550	750.000	-3.5

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: RTX-CLP2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/11/2025 02/11/2025

Client Sample No.: CCAL03 Date Analyzed: 02/14/2025

Lab Sample No.: HSTDCCC750 Data File : PS029179.D Time Analyzed: 03:00

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
2,4,5-TP(Silvex)	9.792	9.696	9.896	707.760	712.500	-0.7
2,4-D	8.894	8.797	8.997	677.370	705.000	-3.9
2,4-DCAA	7.662	7.564	7.764	712.640	750.000	-5.0