

Surrogate Summary

SDG No.: Q1352

Client: Weston Solutions

Analytical Method: 8151A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PS029117.D	PIBLK-PS029117.D	2,4-DCAA	1	500	477	95		32	138
		2,4-DCAA	2	500	487	97		32	138
I.BLK-PS029159.D	PIBLK-PS029159.D	2,4-DCAA	1	500	510	102		32	138
		2,4-DCAA	2	500	498	100		32	138
PB166713BL	PB166713BL	2,4-DCAA	1	500	458	92		32	138
		2,4-DCAA	2	500	429	86		32	138
PB166713BS	PB166713BS	2,4-DCAA	1	500	516	103		32	138
		2,4-DCAA	2	500	483	97		32	138
PB166700TB	PB166700TB	2,4-DCAA	1	500	504	101		32	138
		2,4-DCAA	2	500	375	75		32	138
Q1352-02	TAP-IDW-SOIL-021025	2,4-DCAA	1	500	521	104		32	138
		2,4-DCAA	2	500	372	74		32	138
I.BLK-PS029166.D	PIBLK-PS029166.D	2,4-DCAA	1	500	503	101		32	138
		2,4-DCAA	2	500	496	99		32	138
Q1356-04MS	CARBON-WATERMS	2,4-DCAA	1	500	497	99		32	138
		2,4-DCAA	2	500	470	94		32	138
Q1356-04MSD	CARBON-WATERMSD	2,4-DCAA	1	500	485	97		32	138
		2,4-DCAA	2	500	469	94		32	138
I.BLK-PS029178.D	PIBLK-PS029178.D	2,4-DCAA	1	500	502	100		32	138
		2,4-DCAA	2	500	499	100		32	138