

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 02/14/2025

Lab Sample No.: PSTDCCC050 Data File : PL094190.D Time Analyzed: 13:05

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.058	8.953	9.153	52.300	50.000	4.6
Endrin	6.576	6.472	6.672	49.200	50.000	-1.6
gamma-BHC (Lindane)	4.332	4.227	4.427	50.460	50.000	0.9
Heptachlor	4.920	4.814	5.014	52.730	50.000	5.5
Heptachlor epoxide	5.688	5.582	5.782	52.010	50.000	4.0
Methoxychlor	7.503	7.398	7.598	56.630	50.000	13.3
Tetrachloro-m-xylene	3.544	3.439	3.639	49.550	50.000	-0.9

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL01 Date Analyzed: 02/14/2025

Lab Sample No.: PSTDCCC050 Data File : PL094190.D Time Analyzed: 13:05

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.909	7.810	8.010	55.690	50.000	11.4
Endrin	5.636	5.536	5.736	54.270	50.000	8.5
gamma-BHC (Lindane)	3.605	3.507	3.707	51.600	50.000	3.2
Heptachlor	3.944	3.845	4.045	52.690	50.000	5.4
Heptachlor epoxide	4.726	4.627	4.827	52.360	50.000	4.7
Methoxychlor	6.609	6.509	6.709	58.520	50.000	17.0
Tetrachloro-m-xylene	2.773	2.674	2.874	50.460	50.000	0.9