

PESTICIDE CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL093726.D Date Analyzed: 01/21/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:30

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	9.052	8.950	9.150	18.070	20.000	-9.7
Tetrachloro-m-xylene	3.538	3.490	3.590	18.530	20.000	-7.4
alpha-BHC	3.994	3.940	4.040	9.490	10.000	-5.1
beta-BHC	4.525	4.470	4.580	9.790	10.000	-2.1
gamma-BHC (Lindane)	4.326	4.280	4.380	9.300	10.000	-7.0
Endrin	6.572	6.500	6.640	41.270	50.000	-17.5
4,4'-DDT	7.022	6.950	7.090	82.410	100.000	-17.6
Methoxychlor	7.498	7.430	7.570	190.380	250.000	-23.8

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No. (PEM): PEM - PL093726.D Date Analyzed: 01/21/2025

Lab Sample No.(PEM): PEM Time Analyzed: 10:30

PEM COMPOUND	RT	RT WINDOW		CALC	NOM	%D
		FROM	TO	AMOUNT(ng)	AMOUNT(ng)	
Decachlorobiphenyl	7.909	7.810	8.010	17.950	20.000	-10.3
Tetrachloro-m-xylene	2.775	2.720	2.830	17.900	20.000	-10.5
alpha-BHC	3.277	3.230	3.330	8.620	10.000	-13.8
beta-BHC	3.907	3.860	3.960	9.800	10.000	-2.0
gamma-BHC (Lindane)	3.607	3.560	3.660	8.300	10.000	-17.0
Endrin	5.636	5.570	5.710	42.700	50.000	-14.6
4,4'-DDT	6.034	5.960	6.100	96.510	100.000	-3.5
Methoxychlor	6.609	6.540	6.680	209.940	250.000	-16.0