

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1352

Client: Weston Solutions

Analytical Method: **8081B** **Datafile :** PL094221.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD		Limits	
								Qual	Low	High	RPD
PB166712BS	gamma-BHC (Lindane)	0.5	0.44	ug/L	89				59	134	
	Heptachlor	0.5	0.48	ug/L	97				54	130	
	Heptachlor epoxide	0.5	0.47	ug/L	93				61	133	
	Endrin	0.5	0.48	ug/L	96				60	138	
	Methoxychlor	0.5	0.52	ug/L	104				54	145	