

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

Continuing Calib Date: 02/17/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 15:35 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	9.05	9.05	8.95	9.15	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
gamma-BHC (Lindane)	4.33	4.33	4.23	4.43	0.01
Heptachlor	4.91	4.91	4.81	5.01	0.00
Heptachlor epoxide	5.68	5.68	5.58	5.78	0.00
Endrin	6.57	6.57	6.47	6.67	0.00
Methoxychlor	7.50	7.50	7.40	7.60	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

Continuing Calib Date: 02/17/2025 Initial Calibration Date(s): 01/21/2025 01/21/2025

Continuing Calib Time: 15:35 Initial Calibration Time(s): 10:57 11:51

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Decachlorobiphenyl	7.91	7.91	7.81	8.01	0.00
Tetrachloro-m-xylene	2.77	2.77	2.67	2.87	0.00
gamma-BHC (Lindane)	3.60	3.61	3.51	3.71	0.01
Heptachlor	3.94	3.95	3.85	4.05	0.01
Heptachlor epoxide	4.72	4.73	4.63	4.83	0.01
Endrin	5.63	5.64	5.54	5.74	0.01
Methoxychlor	6.61	6.61	6.51	6.71	0.00