

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 02/14/2025

Lab Sample No.: PSTDCCC050 Data File : PL094202.D Time Analyzed: 16:49

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.051	8.953	9.153	52.560	50.000	5.1
Endrin	6.570	6.472	6.672	51.500	50.000	3.0
gamma-BHC (Lindane)	4.326	4.227	4.427	51.480	50.000	3.0
Heptachlor	4.914	4.814	5.014	52.980	50.000	6.0
Heptachlor epoxide	5.681	5.582	5.782	53.120	50.000	6.2
Methoxychlor	7.497	7.398	7.598	58.130	50.000	16.3
Tetrachloro-m-xylene	3.538	3.439	3.639	50.090	50.000	0.2

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL02 Date Analyzed: 02/14/2025

Lab Sample No.: PSTDCCC050 Data File : PL094202.D Time Analyzed: 16:49

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.905	7.810	8.010	55.310	50.000	10.6
Endrin	5.634	5.536	5.736	55.140	50.000	10.3
gamma-BHC (Lindane)	3.605	3.507	3.707	53.290	50.000	6.6
Heptachlor	3.942	3.845	4.045	53.650	50.000	7.3
Heptachlor epoxide	4.724	4.627	4.827	53.270	50.000	6.5
Methoxychlor	6.606	6.509	6.709	58.410	50.000	16.8
Tetrachloro-m-xylene	2.772	2.674	2.874	51.590	50.000	3.2